

# Declaration of conformity

## Batch

Declaration of conformity given by the supplier of product according to the standard EN ISO IC 17050-1

**WERTLAB**  
**GERMANY**

Establishment applied and using system of quality  
according to the DIN EN ISO 9001: 2015, certificate made out by TÜV Rheinland FRG


Sort of product:	<b>1605 / AS</b>	<b>GRADUATED PIPETTE</b>
Nominal volume:	<b>5</b>	<b>ml</b>
Division:	<b>0,05</b>	<b>ml</b>
Permitted volume error:	<b>± 0,03</b>	<b>ml</b>
Average value of volume:		<b>ml</b>
Standard deviation of measurement:	<b>±</b>	<b>ml</b>
Number of charge:	<b>00 . 00</b>	

Testing instrument: balance – KERN ABT 220-4M; id. number: 34/07;,  
thermometer – enclosed-scale, (0 – 50) °C; 0,01 °C; id. number: 12/12.  
Used standards are in accordance to standards of State Metrology Institute.

We are declaring to our exclusive responsibility, that product, which relates to this declaration responds to the parameters of standard:  
EN ISO 835:2007.

Date of edition: 00. 00. 0000

Confirmed by:

  
Manager of Quality

000000 / 0